

**Konformitätserklärung
Declaration of Conformity**

Wir

**B. Braun Melsungen AG
Carl-Braun-Straße 1
34212 Melsungen
Deutschland/Germany**

We

erklären in eigener Verantwortung,
dass das/die Produkt/e**Peritofix®**

Besteck zur diagnostischen Peritoneallavage

(Artikelnummern siehe Anlage I)

mit den Anforderungen der folgenden Richtlinie
übereinstimmt/übereinstimmenRichtlinie 93/42/EWG des Rates vom 14. Juni
1993über Medizinprodukte
geändert durch Richtlinie 2007/47/EG**Konformitätsbewertungsverfahren**
nach Anhang II (ausgenommen Abschnitt 4)
der oben genannten Richtlinie**Klassifizierung**
gemäß Anhang IX der
oben genannten Richtlinie
Klasse IIa**Benannte Stelle**
TÜV SÜD Product Service GmbH
Ridlerstraße 65
80339 München
Deutschland
Kennnummer 0123**Datum der ersten CE-Kennzeichnung**
1994-12**Gültig bis**
2024-05-26hereby declare in our own responsibility
that the product/s**Peritofix®**

Kit for diagnostic peritoneal lavage

(article numbers see attachment I)

is/are in compliance with the following directive

Council Directive 93/42/EEC of 14th June
1993concerning Medical Devices
amended by Directive 2007/47/EC**Conformity Assessment Procedure**
according to annex II (excluding section 4)
of the Council Directive named above**Classification**
according to annex IX of the
Council Directive named above
Class IIa**Notified Body**
TÜV SÜD Product Service GmbH
Ridlerstraße 65
80339 München
Germany
Identification number 0123**Date of first CE-marking**
1994-12**Valid until**
2024-05-26

Effective

Anlage I / Attachment I

Art.-Nr. / Art. No.	Produktname / Product name	Klasse / Class
4463005	Peritofix [®]	Ila

Amendment Information

Version	Description of the changes
08	Update due to Expiry Date Add Address of the Notified Body

Title: Declaration of Conformity - 39.05.420 - Peritofix Initiator: Caroline ? Herbst

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Herbst, Caroline (fuehcade)
Title: Administrator Regulatory Affairs
Date: Monday, 01 July 2019, 12:13 W. Europe Daylight Time
Meaning: Document signed as Author
=====

UserName: Seidel, Stefan (seidstde)
Title: HC-RA-DE08E - Head of Regulatory Affairs CoE IV Systems
Date: Monday, 01 July 2019, 18:50 W. Europe Daylight Time
Meaning: Approve Document
=====

UserName: Brand, Thomas (brantode)
Title: HC-QM-DE08 Vice President QM for non-active Medical Devices
Date: Monday, 01 July 2019, 21:46 W. Europe Daylight Time
Meaning: Approve Document
=====